

MOUSE INTERLEUKIN-2 ASSAY LAYOUT SHEET
for use with Assay Designs' Catalog No. 900-042 (formerly 90042)

• **TABLE FOR MAKING STANDARDS 1-8:**

Std.	Diluent Vol. (µL)	Vol. Added (µL)	mouse IL-2 Conc. (pg/mL)
1	900	100, Stock	1000
2	500	500, Std. 1	500
3	500	500, Std. 2	250
4	500	500, Std. 3	125
5	500	500, Std. 4	62.5
6	500	500, Std. 5	31.25
7	500	500, Std. 6	15.63
8	500	500, Std. 7	7.81

• **ASSAY PROTOCOL FLOW CHART:**

Well ID:	Blank A1, B1	S0 C1, D1	Standards E1-D3	Samples E3-H12
Standard Diluent	---	50µL	---	---
Std. and/or Sample	---	---	50µL	50µL
Tap plate gently to mix				
Incub. 2 hours @ RT, sealed, shaking				
Asp. & Wash 4 x 400µL				
Antibody	---	50µL	50µL	50µL
Incub. 1 hour @ RT, sealed, shaking				
Asp. & Wash 4 x 400µL				
Conjugate	---	50µL	50µL	50µL
Incub. 30 min. @ RT, sealed, shaking				
Asp. & Wash 4 x 400µL				
TMB Substrate	50µL	50µL	50µL	50µL
Incub. 15 min. @ RT, shaking				
Stop Solution (1N HCl)	50µL	50µL	50µL	50µL
Read OD at 450nm				

MOUSE IL-2 PLATE LAYOUT:

A1 Blank	A2 Std 3	A3 Std 7	A4	A5	A6	A7	A8	A9	A10	A11	A12
B1 Blank	B2 Std 3	B3 Std 7	B4	B5	B6	B7	B8	B9	B10	B11	B12
C1 S0	C2 Std 4	C3 Std 8	C4	C5	C6	C7	C8	C9	C10	C11	C12
D1 S0	D2 Std 4	D3 Std 8	D4	D5	D6	D7	D8	D9	D10	D11	D12
E1 Std 1	E2 Std 5	E3	E4	E5	E6	E7	E8	E9	E10	E11	E12
F1 Std 1	F2 Std 5	F3	F4	F5	F6	F7	F8	F9	F10	F11	F12
G1 Std 2	G2 Std 6	G3	G4	G5	G6	G7	G8	G9	G10	G11	G12
H1 Std 2	H2 Std 6	H3	H4	H5	H6	H7	H8	H9	H10	H11	H12

Kit Lot No. _____ Exp. Date _____ Date _____ Tech. _____

1st Incub.: Start Time _____ Temp. _____
 End Time _____ Temp. _____

3rd Incub.: Start Time _____ Temp. _____
 End Time _____ Temp. _____

2nd Incub.: Start Time _____ Temp. _____
 End Time _____ Temp. _____

4th Incub.: Start Time _____ Temp. _____
 End Time _____ Temp. _____